



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Onset Computer Corporation
470 MacArthur Blvd.
Bourne, MA 02532

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President
Expiry Date: 18 April 2026
Certificate Number: AC-3091



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Onset Computer Corporation

470 MacArthur Blvd.
Bourne, MA 02532
Justin Payne 508-743-3379

CALIBRATION

Valid to: **April 18, 2026**

Certificate Number: **AC-3091**

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature Dataloggers	(0 to 120) °C	0.13 °C	Baths, Temperature Indicator w/ Probe
Cryogenic Temperature Dataloggers	(-70 to 0) °C	0.5 °C	Baths, Temperature Indicator w/ Probe
Humidity Dataloggers	(25 to 90) %RH	1.3 %RH	Comparison to Chilled Mirror, Environmental Chamber

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-3091.



Jason Stine, Vice President