



# CERTIFICATE OF ACCREDITATION

## The ANSI National Accreditation Board

Hereby attests that

**Onset Computer Corporation**  
470 MacArthur Blvd.  
Bourne, MA 02532

Fulfills the requirements of

**ISO/IEC 17025:2017**

In the field of

**CALIBRATION**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

Jason Stine, Vice President

Expiry Date: 18 April 2028

Certificate Number: AC-3091



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**Onset Computer Corporation**

470 MacArthur Blvd.  
 Bourne, MA 02532  
 Justin Payne 508-743-3379

**CALIBRATION**

ISO/IEC 17025 Accreditation Granted: **17 April 2026**

Certificate Number: **AC-3091**

Certificate Expiry Date: **18 April 2028**


**Thermodynamic**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature Dataloggers	(0 to 120) °C	0.13 °C	Comparison to Baths, Temperature Indicator with Probe
Cryogenic Temperature Dataloggers	(-70 to 0) °C	0.5 °C	Comparison to Baths, Temperature Indicator with Probe
Humidity Dataloggers	(25 to 90) %RH	1.3 %RH	Comparison to Chilled Mirror, Environmental Chamber

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. Unless otherwise specified in the far-right column, the calibration procedure utilized by the laboratory was internally developed and validated..



Jason Stine, Vice President